



DOC-1138

Declaration of Conformity

Radio Equipment Compliance

RRR Manufacturing Pty Ltd
2-6 Skinner Avenue
Riverwood, NSW 2210
Australia

Contents

Part 1 European Commission:	3
Part 2 United Kingdom:	4

Manufacturer	RRR Manufacturing Pty Ltd 2-6 Skinner Avenue Riverwood, NSW 2210 Australia +61-1300-859-903
Product Description	Automated External Defibrillator
Product Name	CellAED®

All supporting documentation is retained under the premises of the manufacturer.

Part 1 European Commission:

The object is in conformity with the relevant statutory requirements:

Radio Equipment Directive (RED) 2014/53/EU

European Representative	mdi Europa GmbH Langenhagener Straße 71 D-30855 Langenhagen Germany		
Notified Body	Phoenix Testlab GmbH Königswinkel 10, D-32825 Blomberg, Germany Notified Body Number: 0700		
Classification	<input checked="" type="checkbox"/> Equipment Class 1	<input type="checkbox"/> Equipment Class 2	
EC Certificate number	20-211357		
Scope Application	All batches		
Software	Description/Model:	CellAED® 2.2	
	Version:	1.0.X	
Battery Information	Model:	CP155070-4S (60Cell)	
	Battery features:	Li-MnO2	
	V. Nominal:	12V	
	Rated Capacity:	900 mAh	
Wireless Module	Description:	SARA-U201	
	Module type:	GSM/UMTS transceiver	
	Certificate number:	ATCB023842 issue 7	

Standards Applied	
<input checked="" type="checkbox"/> Article 3.1(a)	EN 50566:2017
	IEC 60601-1:2014
	EN 62209-2:2010
	EN 62311:2008
	EN 62368-1:2014
<input checked="" type="checkbox"/> Article 3.1(b)	EN 60601-1-2:2015
	EN 301 489-1: v2.2.3 (2019-11)
	EN 301 489-52: v1.1.0 (2016-11)
<input checked="" type="checkbox"/> Article 3.2	EN 301 511: v12.5.1 (2017-03)
	EN 301 908-1: v11.1.1 (2016-07) & v13.1.1 (2019-11)
	EN 301 908-2: v11.1.2 (2017-08)

Part 2 United Kingdom:

The object is in conformity with the relevant statutory requirements:

UK Radio Equipment Regulations 2017

UK Responsible Person	Psephos Limited Sussex Innovation Centre, Science Park Square, Brighton, East Sussex, United Kingdom +44-1273-704-527		
Classification	<input checked="" type="checkbox"/>	Equipment Class 1	<input type="checkbox"/> Equipment Class 2
Scope Application	All batches		
Software	Description/Model:	CellAED® 2.2	
	Version:	1.0.X	
Battery Information	Model:	CP155070-4S (60Cell)	
	Battery features:	Li-MnO2	
	V. Nominal:	12V	
	Rated Capacity:	900 mAh	
Wireless Module	Description:	SARA-U201	
	Module type:	GSM/UMTS transceiver	
	Certificate number:	ATCB023842 issue 7	

Standards Applied			
<input checked="" type="checkbox"/> Article 6-(1) (a)	EN 50566:2017		
	IEC 60601-1:2014		
	EN 62209-2:2010		
	EN 62311:2008		
	EN 62368-1:2014		
<input checked="" type="checkbox"/> Article 6-(1) (b)	EN 60601-1-2:2015		
	EN 301 489-1: v2.2.3 (2019-11)		
	EN 301 489-52: v1.1.0 (2016-11)		
<input checked="" type="checkbox"/> Article 6-(2)	EN 301 511: v12.5.1 (2017-03)		
	EN 301 908-1: v11.1.1 (2016-07) & v13.1.1 (2019-11)		
	EN 301 908-2: v11.1.2 (2017-08)		
Conformity Assessment Body	N/A		
Assessment Body No.	N/A		
Type Examination No.	N/A		
<input checked="" type="checkbox"/>	Equipment Class 1	<input type="checkbox"/>	Equipment Class 2



DOC-1138 Declaration of Conformity Radio Equipment Compliance

Signed for on behalf of RRR Manufacturing Pty Ltd who are responsible for the issuance, maintenance, extension, suspension or withdrawal of this Declaration of Conformity.

A handwritten signature in blue ink, appearing to read "Trudap".

Chief Compliance Officer

22 – 09 – 2021

Date